

Product Information

Product Name	TE06
Alias	I-6
Lot Number	TE06-DL-01
Parent Material	TE06-MCB-01
Depositor	Technion
Banked by	WiCell
Thaw Recommendation	Thaw 1 vial into 1 well of a 6 well plate.
Culture Platform	Feeder Dependent
	Medium: hES Medium
	Matrix: MEF
Protocol	WiCell Feeder Dependent Protocol
Passage Number	p50 These cells were cultured for 49 passages prior to freeze. WiCell adds +1 to the passage number at freeze so that the number on the vial best represents the overall passage number of the cells at thaw.
Date Viald	16-September-2009
Vial Label	TE06-DL-01 p50 LD 16 SEPT 2009 SOPCC035D
Biosafety and Use Information	Appropriate biosafety precautions should be followed when working with these cells. The end user is responsible for ensuring that the cells are handled and stored in an appropriate manner. WiCell is not responsible for damages or injuries that may result from the use of these cells. Cells distributed by WiCell are intended for research purposes only and are not intended for use in humans.

Testing Performed by WiCell

Test Description	Test Provider	Test Method	Test Specification	Result
Post-Thaw Viable Cell Recovery	WiCell	SOP-CH-305	≥ 15 Undifferentiated Colonies, ≤ 30% Differentiation	Pass
Identity by STR	UW Molecular Diagnostics Laboratory	PowerPlex 1.2 System by Promega	Positive Identity ¹	Pass
	¹ An anomalous band pattern has been observed in this TE06 cell lot for STR. See additional information regarding similar anomalies at: "A Genetic Basis for Anomalous Band Patterns Encountered During DNA STR Profiling", Clayton, T.M., et al. J. Forensic Sci, Nov. 2004, Vol. 49, No. 6.			
Sterility - Direct transfer method	Apptec	30744	No contamination detected	Pass
Mycoplasma	Bionique	M250	No contamination detected	Pass
Karyotype by G-banding	WiCell	SOP-CH-003	Normal karyotype	Pass
Flow Cytometry for ESC Marker Expression	UW Flow Cytometry Laboratory	SOP-CH-101 SOP-CH-102 SOP-CH-103 SOP-CH-105	Report - no specification	See report



Product Information and Testing - Amended

Amendment(s):

Reason for Amendment	Date
CoA updated to include copyright information.	See signature
CoA updated for format changes, including adding fields of thaw recommendation, vial label, protocol, and banked by, and removal of footnotes.	28-JUN-2013
CoA updated for format changes, clarification of test specifications, test method, addition of test provider, culture platform, and electronic signature, and reference to WiCell instead of the NSCB	17-AUG-2010
Original CoA	01-APR-2010

Date of Lot Release	Quality Assurance Approval
01-April-2010	<p style="text-align: right;">12/31/2013</p> <p style="text-align: center;">X AMC</p> <p>AMC Quality Assurance Signed by: XXXXXXXXXX</p>

Short Tandem Repeat Analysis*

Sample Report: 4126-STR

UW HLA#: 62408

Sample Date: 01/25/10

Received Date: 01/25/10

Requestor: WiCell Research Institute

Test Date: 01/26/10

File Name: 100127, 100129 Report Date: 02/02/10

Sample Name: (label on tube) 4126-STR

Description: DNA Extracted by WiCell
253.44ug/mL; 260/280 = 1.93

Locus	Repeat #	STR Genotype
D16S539	5, 8-15	9,13
D7S820	6-14	8,10
D13S317	7-15	8,11
D5S818	7-15	12,12
CSF1PO	6-15	11,11
TPOX	6-13	10,11
Amelogenin	NA	X,Y
TH01	5-11	8,9.3
vWA	11, 13-21	16,17

Comments: Based on the DNA 4126-STR dated and received on 01/25/10 from WI Cell, this sample (UW HLA# 62408) generally matches the STR profile of the human stem cell line TE06 comprising 14 allelic polymorphisms across the 8 STR loci analyzed with the exception that the 4126-STR DNA sample displays a homozygous 11,11 genotype at the CSF1PO loci rather than a heterozygous 10,11 genotype that is published for TE06. Other than this discrepancy noted at the CSF1PO loci, no STR polymorphisms other than those corresponding to the human TE06 stem cell line were detected and the concentration of DNA required to achieve an acceptable STR genotype (signal/noise) was equivalent to that required for the standard procedure (~1 ng/amplification reaction) from human genomic DNA. These results suggest that the 4126-STR DNA sample submitted corresponds to the TE06 stem cell line with a discrepancy at the CSF1PO loci and that it was not contaminated with any other human stem cells or a significant amount of mouse feeder layer cells. Sensitivity limits for detection of STR polymorphisms unique to either this or other human stem cell lines is estimated to be ~5%.

A

HLA/Molecular Diagnostics Laboratory

HLA/Molecular Diagnostics Laboratory

* Testing to assess engraftment following bone marrow transplantation was accomplished by analysis of human genetic polymorphisms at STR loci. This methodology has not yet been approved by the FDA and is for investigational use only.

Test Facility:

This report is confidential. No part may be used for advertising or public announcement without written permission. Results apply only to the sample(s) tested.

WiCell Research Institute

Report Number
830702
Page 1 of 1

March 16, 2010
P.O. #:

STERILITY TEST REPORT

Sample Information: hES Cells
1: TE06-DL-01, # 3979
2: SA01-DL-03, # 8903
3: WA07-FTDL-03, # 7194

Date Received: February 25, 2010
Date in Test: March 01, 2010
Date Completed: March 15, 2010

Test Information: Test Codes: 30744, 30744A
Immersion, USP / 21 CFR 610.12
Procedure #: BS210WCR.201

TEST PARAMETERS	PRODUCT	
	Approximate Volume Tested	0.5 mL
Number Tested	6	6
Type of Media	SCD	FTM
Media Volume	400 mL	400 mL
Incubation Period	14 Days	14 Days
Incubation Temperature	20 °C to 25 °C	30 °C to 35 °C
RESULTS	6 NEGATIVE	6 NEGATIVE

QA Reviewer

Date

Technical Reviewer

Date

Testing conducted in accordance with current Good Manufacturing Practices.



APPENDIX

BIONIQUÉ® TESTING LABORATORIES, INC.

Document ID #: DCF9002E
Title: QUALITY ASSURANCE REPORT - GMP
Effective Date: 01/04/10
Edition #: 02

QUALITY ASSURANCE REPORT – GMP

<u>TEST PERFORMED</u>	<u>PROCEDURAL REFERENCE</u>	<u>TEST PERFORMED</u>	<u>PROCEDURAL REFERENCE</u>
<input checked="" type="checkbox"/> M-250	SOP's 3008, 3011, 3013	<input type="checkbox"/> M-700	SOP's 3008, 3009, 3010
<input type="checkbox"/> M-300	SOP's 3008, 3014	<input type="checkbox"/> M-800	SOP's 3008, 3011, 3016
<input type="checkbox"/> M-350	SOP's 3008, 3014, 3015		

Bionique Sample ID #(s) 59959 59960

This testing procedure was performed in compliance with the FDA's Current Good Manufacturing Practice (cGMP) standards (to the extent that the regulations pertain to the procedures performed) as specified in the Code of Federal Regulations, Title 21 Parts 210 and 211 [21 CFR 210 & 211]. All related records derived from the test procedures have been reviewed by the Quality Assurance Department. The individual's signature below verifies that the methods and procedures referenced above have been followed and that the Final Report accurately reflects the raw data generated during the course of the procedures. All records, including raw data and final reports are archived on site for a minimum of seven years.

The specified test's procedures determine the intervals at which samples are inspected. The medium used for testing must pass quality control mycoplasmal growth promotion testing and sterility testing. Traceability of all of the components used is assured and supporting documentation can be supplied upon request.

Quality Assurance Review Date: 2/11/10

Reviewed By QA Assistant: _____

NOTE:

1. Prior to receipt at Bionique® Testing Laboratories, Inc., the stability of the test article is the responsibility of the company submitting the sample. Bionique Testing Laboratories Inc. will assume responsibility for sample stability following receipt and prior to being placed on test.
2. This test is for the detection of microbiological growth and does not require statistical validation.

Document ID #: DCF9002E
Title: QUALITY ASSURANCE REPORT - GMP
Effective Date: 05/21/09
Edition #: 02

REFERENCES

Regulatory:

1. Department of Health and Human Services, Food and Drug Administration (USA) [FDA]. Code of Federal Regulations [CFR], Title 21 CFR Part 210, Current Good Manufacturing Practice in Manufacturing, Processing, Packing, or Holding of Drugs; General. FDA. Office of the Federal Register, National Archives and Records Department.
2. Department of Health and Human Services, Food and Drug Administration (USA) [FDA]. Code of Federal Regulations [CFR], Title 21 CFR Part 211, Current Good Manufacturing Practice for Finished Pharmaceuticals. FDA. Office of the Federal Register, National Archives and Records Department.
3. Department of Health and Human Services, Food and Drug Administration (USA) [FDA]. Points to Consider in the Characterization of Cell Lines Used to Produce Biologicals, Director, Center for Biologics Evaluation and Research, FDA. May, 1993. Docket No. 84N-0154.
4. Department of Health and Human Services, Food and Drug Administration (USA) [FDA]. Code of Federal Regulations [CFR], Title 21 CFR Part 610.30, General Biological Products Standards; Subpart D, Test for Mycoplasma. FDA. Office of the Federal Register, National Archives and Records Department.

General:

1. Barile MF, Kern J. Isolation of Mycoplasma arginini from commercial bovine sera and its implication in contaminated cell cultures. Proceedings of the Society for Experimental Biology and Medicine, Volume 138, Number 2, November 1971.
2. Chen, T.R. In situ detection of mycoplasma contamination in cell cultures by fluorescent Hoechst 33258 stain. Experimental Cell Research, 104: 255-262, 1977.
3. Carolyn K. Lincoln and Daniel J. Lundin. Mycoplasma Detection and Control. U. S. Fed. for Culture Collections Newsletter, Vol. 20, Number 4, 1990.
4. Fetal Bovine Serum; Proposed Guideline. National Committee For Clinical Laboratory Standards (NCCLS), Vol. 10, Number 6, 1990. (NCCLS publication M25-P).
5. McGarrity GJ, Sarama J, Vanaman V. Cell Culture Techniques. ASM News, Vol. 51, No. 4, 1985.
6. Tully JG, Razin S. Methods in Mycoplasma, Volumes I and II. Academic Press, N.Y., 1983.
7. Barile MF, Razin S, Tully JG, Whitcomb RF. The Mycoplasmas, Volumes 1-4. Academic Press, N.Y., 1979.
8. <http://www.bionique.com/> - Safe Cells Insights



APPENDIX IV

Page 1 of 2

Document#: DCF3013D
Edition#: 10
Effective Date: 07/15/2003
Title: **M-250 FINAL REPORT SHEET**

M-250 FINAL REPORT

Direct Specimen Culture
Procedure 3008, 3011, 3013

TO: **Wicell QA**

BTL SAMPLE ID#: **59959** P.O.#: DATE REC'D: **01/14/2010**

TEST/CONTROL ARTICLE:

TE06-DL-1-Mp62 #4126

LOT#: **NA**

DIRECT CULTURE SET-UP (DAY 0)

DATE: **01/14/2010**

INDICATOR CELL LINE (VERO)

SEE DNA FLUOROCHROME RECORD SHEET

				DATE
THIOGLYCOLLATE BROTH	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>
BROTH-FORTIFIED COMMERCIAL <u>0.5</u> mL SAMPLE	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>
<u>6.0</u> mL BROTH	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>
BROTH-MODIFIED HAYFLICK <u>0.5</u> mL SAMPLE	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>
<u>6.0</u> mL BROTH	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>
BROTH-HEART INFUSION <u>0.5</u> mL SAMPLE	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>
<u>6.0</u> mL BROTH	DAY 7	+	⊖	<u>01/21/2010</u>
	DAY 28	+	⊖	<u>02/11/2010</u>

(See Reverse)

Document#: DCF3013D
 Edition#: 10
 Effective Date: 07/15/2003
 Title: M-250 FINAL REPORT SHEET

SAMPLE ID#:	59959	AEROBIC	MICROAEROPHILIC	DATE
AGAR PLATES-FORTIFIED COMMERCIAL	DAY 7	+ ⊖	+ ⊖	<u>01/21/2010</u>
	DAY 14	+ ⊖	+ ⊖	<u>01/28/2010</u>
	DAY 21	+ ⊖	+ ⊖	<u>02/04/2010</u>
AGAR PLATES-MODIFIED HAYFLICK	DAY 7	+ ⊖	+ ⊖	<u>01/21/2010</u>
	DAY 14	+ ⊖	+ ⊖	<u>01/28/2010</u>
	DAY 21	+ ⊖	+ ⊖	<u>02/04/2010</u>
AGAR PLATES-HEART INFUSION	DAY 7	+ ⊖	+ ⊖	<u>01/21/2010</u>
	DAY 14	+ ⊖	+ ⊖	<u>01/28/2010</u>
	DAY 21	+ ⊖	+ ⊖	<u>02/04/2010</u>

BROTH SUBCULTURES (DAY 7)DATE: 01/21/2010

AGAR PLATES-FORTIFIED COMMERCIAL	DAY 7	+ ⊖	+ ⊖	<u>01/28/2010</u>
	DAY 14	+ ⊖	+ ⊖	<u>02/04/2010</u>
	DAY 21	+ ⊖	+ ⊖	<u>02/11/2010</u>
AGAR PLATES-MODIFIED HAYFLICK	DAY 7	+ ⊖	+ ⊖	<u>01/28/2010</u>
	DAY 14	+ ⊖	+ ⊖	<u>02/04/2010</u>
	DAY 21	+ ⊖	+ ⊖	<u>02/11/2010</u>
AGAR PLATES-HEART INFUSION	DAY 7	+ ⊖	+ ⊖	<u>01/28/2010</u>
	DAY 14	+ ⊖	+ ⊖	<u>02/04/2010</u>
	DAY 21	+ ⊖	+ ⊖	<u>02/11/2010</u>

RESULTS: No detectable mycoplasmal contamination

2/11/10
Date

M-250 Procedural Summary: The objective of this test is to ascertain whether or not detectable mycoplasmas are present in an in vitro cell culture sample, be it a primary culture, hybridoma, master seed stock or cell line. This procedure combines an indirect DNA staining approach to detect non-cultivable mycoplasmas with a direct culture methodology utilizing three different mycoplasmal media formulations. The indirect approach involves the inoculation of the sample into a mycoplasma-free VERO (ATCC) indicator cell line and performing a DNA fluorochrome assay after 72-120 hours of incubation. The direct culture aspect of the test utilizes three different mycoplasmal media including both broth and agar formulations. The sample is inoculated into each of the 3 broth formulations and also onto duplicate plates (0.1 mL/plate) for each of the 3 agar formulations. Subculture from broth to fresh agar plates is carried out after 7 days incubation. Agar plates are incubated aerobically and microaerophilically in order to detect any colony forming units morphologically indicative of mycoplasmal contamination. Issuance of the final report with signature of the Laboratory Director signifies that the required controls were performed concurrently with the test sample(s) as detailed in the referenced SOPs and that all test conditions have been found to meet the required acceptance criteria for a valid test, including the appropriate results for the positive and negative controls.



APPENDIX I

Document #: DCF3008A
Edition #: 06
Effective date: 9/17/2003
Title: DNA FLUOROCHROME ASSAY RESULTS

DNA-FLUOROCHROME ASSAY RESULTS
Procedures 3008, 3009, 3011

Sample ID # 59959 M-250 Date Rec'd: 01/14/2010 P.O. #

Indicator Cells Inoculated: Date/Initials: 1/14/10 / JA

Fixation: Date/Initials: 1/18/10 / KG

Staining: Date/Initials: 1/18/10 / KG

TEST/CONTROL ARTICLE:

TE06-DL-1-Mp62 #4126

LOT# NA

Wicell QA

DNA FLUOROCHROME ASSAY RESULTS:

X NEGATIVE: A reaction with staining limited to the nuclear region, which indicates no mycoplasmal contamination.

POSITIVE: A significant amount of extranuclear staining which strongly suggests mycoplasmal contamination.

INCONCLUSIVE: A significant amount of extranuclear staining consistent with low - level mycoplasmal contamination or nuclear degeneration.

A significant amount of extranuclear staining consistent with bacterial, fungal or other microbial contaminant or viral CPE. Morphology not consistent for mycoplasmal contamination.

COMMENTS:

Date: 1/18/10 Results Read by: KG Date of Review: 1/18/10 Reviewed by: SWA

Report Date: December 14, 2009

Case Details:

Cell Line: TE06-DL-01 (1076)

Passage #: 57

Date Completed: 12/14/2009

Cell Line Gender: Male

Investigator:

Specimen: hESC on MEF feeder

Date of Sample: 12/7/2009


Tests, Reason for: DL testing

Results: 46,XY

Completed by CLSp(CG), on 12/14/2009

Reviewed and interpreted by PhD, FACMG, on 12/14/2009

Interpretation: No abnormalities were detected at the stated band level of resolution.



Cell: S01-04

Slide: B-18

Slide Type: Karyotyping

of Cells Counted:

of Cells Karyotyped:

of Cells Analyzed:

Band Level:

Results Transmitted by Fax / Email / Post

Sent By: _____



QC Review By: _____

Date: _____

Sent To: _____

Results Recorded: _____

antigen2:	SSEA4 - antigen2 +	SSEA4 + antigen2 +	SSEA4 + antigen2 -	SSEA4 - antigen2 -	ALL SSEA4 +	ALL antigen2 +
SSEA3	1.56	56.10	33.40	8.91	89.50	57.66
TRA1-60	2.11	85.30	6.80	5.84	92.10	87.41
TRA1-81	1.98	85.70	6.52	5.77	92.22	87.68
Oct-4	1.77	80.50	10.90	6.84	91.40	82.27
SSEA1	1.69	14.10	78.50	5.65	92.60	15.79



* SSEA3 was removed due to a technical error with the staining of SSEA3.