



Certificate of Analysis - Amended Fast Track Distribution Lot

Product Description	TE06
Cell Line Provider	Technion (Israel)
Distribution Lot Number	TE06-FTDL-1
Date Viald	04-May-2009
Passage Number	46
Culture Method	SOP-CC-020C, SOP-CC-030C
Cryopreservation Method	SOP-CC-035D

The following testing specifications have been met for the specified product lot:

Test Description	Test Method	Test Specification	Result
Post-Thaw Viable Cell Recovery	SOP-CH-305C	Viable cells recovered	Pass
Identity by STR	SOP-SS-006A	Positive identity	Pass ¹
Sterility	Apptec Protocol 30744 Rev. 1	No contamination detected	Pass
Mycoplasma	Bionique Method M250	No contamination detected	Pass
Karyotype by G-banding	SOP-CH-003B	Normal karyotype	Pass

Electronic versions of this certificate of analysis (CoA) complete with electronic copies of individual reports, results, and procedures are available on our website, www.wicell.org. There are also archived CoAs for past cell lots.

Cells distributed by the National Stem Cell Bank are intended for research purposes only and are not intended for use in humans. These cells have undergone testing and are not known to harbor pathogens. However, appropriate biosafety precautions should be followed when working with these cells. The end user is responsible for ensuring that the cells are handled and stored in an appropriate manner. The NSCB is not responsible for damages or injuries that may result from the use of these cells.

Please visit the technical service portion of the website for assistance with your human ES Cells. The knowledgeable technical support staff can assist with embryonic stem cell culture concerns, training, and any other customer service concerns you may encounter.

¹ Identity generally matches the STR profile of the human stem cell line TE06 comprising 14 allelic polymorphisms across the 8 STR loci analyzed with the exception that the DNA displays a homozygous 11, 11 genotype at the CSFIPO loci rather than a heterozygous 10,11 genotype that is published for TE06. No other STR polymorphisms other than those corresponding to the human TE06 stem cell line were detected, demonstrating it was not contaminated with any other human stem cells or a significant amount of mouse feeder layer cells. See the report for this test on the NSCB website for more information.



Certificate of Analysis - Amended Fast Track Distribution Lot

Amendment(s):

Reason for Amendment	Date
CoA updated to include copyright information, electronic signature, and WiCell logo. Links updated.	See signature
Original CoA	23-Jul-2009

Date of Lot Release	Quality Assurance Approval
31-July-2009	<p style="text-align: right;">12/31/2013</p> <p>X AMC</p> <p>AMC Quality Assurance Signed by: [REDACTED]</p>

©2009 WiCell Research Institute

The material provided under this certificate has been subjected to the tests specified and the results and data described herein are accurate based on WiCell's reasonable knowledge and belief. Appropriate Biosafety Level practices and universal precautions should always be used with this material. For clarity, the foregoing is governed solely by WiCell's Terms and Conditions of Service, which can be found at <http://www.wicell.org/privacyandterms>.

Short Tandem Repeat Analysis*

Sample Report: 4873-STR

UW HLA#: 61152

Sample Date: 06/18/09

Received Date: 06/18/09

Requestor: WiCell Research Institute

Test Date: 06/23/09

File Name: 090624

Report Date: 06/25/09

Sample Name: (label on tube) 4873-STR

Description: DNA Extracted by WiCell
234.84 ug/mL; 260/280 = 1.97

Locus	Repeat #	STR Genotype
D16S539	5, 8-15	9,13
D7S820	6-14	8,10
D13S317	7-15	8,11
D5S818	7-15	12,12
CSF1PO	6-15	11,11
TPOX	6-13	10,11
Amelogenin	NA	X,Y
TH01	5-11	8,9.3
vWA	11, 13-21	16,17

Comments: Based on the DNA 4873-STR dated 06/18/09 and received on 06/18/09 from WI Cell, this sample (UW HLA# 61152) generally matches the STR profile of the human stem cell line TE06 comprising 14 allelic polymorphisms across the 8 STR loci analyzed with the exception that the 4873-STR DNA sample displays a homozygous 11,11 genotype at the CSF1PO loci rather than a heterozygous 10,11 genotype that is published for TE06. Other than this discrepancy noted at the CSF1PO loci, no STR polymorphisms other than those corresponding to the human TE06 stem cell line were detected and the concentration of DNA required to achieve an acceptable STR genotype (signal/noise) was equivalent to that required for the standard procedure (~1 ng/amplification reaction) from human genomic DNA. These results suggest that the 4873-STR DNA sample submitted corresponds to the TE06 stem cell line with a discrepancy at the CSF1PO loci and that it was not contaminated with any other human stem cells or a significant amount of mouse feeder layer cells. Sensitivity limits for detection of STR polymorphisms unique to either this or other human stem cell lines is ~5%.

 Manager Date
 HLA/Molecular Diagnostics Laboratory

 PhD, Director Date
 HLA/Molecular Diagnostics Laboratory

* Testing to assess engraftment following bone marrow transplantation was accomplished by analysis of human genetic polymorphisms at STR loci. This methodology has not yet been approved by the FDA and is for investigational use only.

Test Facility:

This report is confidential. No part may be used for advertising or public announcement without written permission. Results apply only to the sample(s) tested.



Report Number
809726
Page 7 of 7

WiCell Research Institute

June 04, 2009
P.O. #:

STERILITY TEST REPORT

Sample Information: hES Cells
6: TE06-FTDL-1 #0158

Date Received: May 19, 2009
Date in Test: May 20, 2009
Date Completed: June 03, 2009

Test Information: Test Codes: 30744, 30744A
Immersion, USP / 21 CFR 610.12
Procedure #: BS210WCR.201

TEST PARAMETERS	PRODUCT	
Approximate Volume Tested	0.5 mL	0.5 mL
Number Tested	2	2
Type of Media	SCD	FTM
Media Volume	400 mL	400 mL
Incubation Period	14 Days	14 Days
Incubation Temperature	20 °C to 25 °C	30 °C to 35 °C
RESULTS	2 NEGATIVE	2 NEGATIVE

Page 1 Signed

QA Reviewer

Date

Page 1 Signed

Technical Reviewer

Date

Testing conducted in accordance with current Good Manufacturing Practices.





APPENDIX IV

Document#: DCF3013D
Edition#: 10
Effective Date: 07/15/2003
Title: M-250 FINAL REPORT SHEET

M-250 FINAL REPORT

Direct Specimen Culture
Procedure 3008, 3011, 3013

TO: Wicell OA

BTL SAMPLE ID#: 57704 P.O.#: DATE REC'D: 06/11/2009

TEST/CONTROL ARTICLE:
TE06-FTDL-01-K #4873

LOT#: NA

DIRECT CULTURE SET-UP (DAY 0) DATE: 06/11/2009

INDICATOR CELL LINE (VERO) SEE DNA FLUOROCHROME RECORD SHEET

DATE

	DAY 7	+	⊖	DATE
THIOGLYCOLLATE BROTH	DAY 7	+	⊖	<u>06/18/2009</u>
	DAY 28	+	⊖	<u>07/09/2009</u>
BROTH-FORTIFIED COMMERCIAL	DAY 7	+	⊖	<u>06/18/2009</u>
<u>0.5</u> mL SAMPLE	DAY 28	+	⊖	<u>07/09/2009</u>
<u>6.0</u> mL BROTH	DAY 7	+	⊖	<u>06/18/2009</u>
BROTH-MODIFIED HAYFLICK	DAY 28	+	⊖	<u>07/09/2009</u>
<u>0.5</u> mL SAMPLE	DAY 7	+	⊖	<u>06/18/2009</u>
<u>6.0</u> mL BROTH	DAY 28	+	⊖	<u>07/09/2009</u>
BROTH-HEART INFUSION	DAY 7	+	⊖	<u>06/18/2009</u>
<u>0.5</u> mL SAMPLE	DAY 28	+	⊖	<u>07/09/2009</u>
<u>6.0</u> mL BROTH				

(See Reverse)

Document#: DCF3013D
 Edition#: 10
 Effective Date: 07/15/2003
 Title: M-250 FINAL REPORT SHEET

SAMPLE ID#:	57704	AEROBIC	MICROAEROPHILIC	DATE
AGAR PLATES-FORTIFIED COMMERCIAL	DAY 7	+ ⊖	+ ⊖	<u>06/18/2009</u>
	DAY 14	+ ⊖	+ ⊖	<u>06/25/2009</u>
	DAY 21	+ ⊖	+ ⊖	<u>07/02/2009</u>
AGAR PLATES-MODIFIED HAYFLICK	DAY 7	+ ⊖	+ ⊖	<u>06/18/2009</u>
	DAY 14	+ ⊖	+ ⊖	<u>06/25/2009</u>
	DAY 21	+ ⊖	+ ⊖	<u>07/02/2009</u>
AGAR PLATES-HEART INFUSION	DAY 7	+ ⊖	+ ⊖	<u>06/18/2009</u>
	DAY 14	+ ⊖	+ ⊖	<u>06/25/2009</u>
	DAY 21	+ ⊖	+ ⊖	<u>07/02/2009</u>

BROTH SUBCULTURES (DAY 7)DATE: 06/18/2009

AGAR PLATES-FORTIFIED COMMERCIAL	DAY 7	+ ⊖	+ ⊖	<u>06/25/2009</u>
	DAY 14	+ ⊖	+ ⊖	<u>07/02/2009</u>
	DAY 21	+ ⊖	+ ⊖	<u>07/09/2009</u>
AGAR PLATES-MODIFIED HAYFLICK	DAY 7	+ ⊖	+ ⊖	<u>06/25/2009</u>
	DAY 14	+ ⊖	+ ⊖	<u>07/02/2009</u>
	DAY 21	+ ⊖	+ ⊖	<u>07/09/2009</u>
AGAR PLATES-HEART INFUSION	DAY 7	+ ⊖	+ ⊖	<u>06/25/2009</u>
	DAY 14	+ ⊖	+ ⊖	<u>07/02/2009</u>
	DAY 21	+ ⊖	+ ⊖	<u>07/09/2009</u>

RESULTS: No detectable mycoplasmal contamination

7-9-09
Date

Laboratory Director

Ph.D.

M-250 Procedural Summary: The objective of this test is to ascertain whether or not detectable mycoplasmas are present in an *in vitro* cell culture sample, be it a primary culture, hybridoma, master seed stock or cell line. This procedure combines an indirect DNA staining approach to detect non-cultivable mycoplasmas with a direct culture methodology utilizing three different mycoplasmal media formulations. The indirect approach involves the inoculation of the sample into a mycoplasma-free VERO (ATCC) indicator cell line and performing a DNA fluorochrome assay after 72-120 hours of incubation. The direct culture aspect of the test utilizes three different mycoplasmal media including both broth and agar formulations. The sample is inoculated into each of the 3 broth formulations and also onto duplicate plates (0.1 mL/plate) for each of the 3 agar formulations. Subculture from broth to fresh agar plates is carried out after 7 days incubation. Agar plates are incubated aerobically and microaerophilically in order to detect any colony forming units morphologically indicative of mycoplasmal contamination. Issuance of the final report with signature of the Laboratory Director signifies that the required controls were performed concurrently with the test sample(s) as detailed in the referenced SOPs and that all test conditions have been found to meet the required acceptance criteria for a valid test, including the appropriate results for the positive and negative controls.



APPENDIX I

Document #: DCF3008A
Edition #: 06
Effective date: 9/17/2003
Title: DNA FLUOROCHROME ASSAY RESULTS

DNA-FLUROCHROME ASSAY RESULTS

Procedures 3008, 3009, 3011

Sample ID # 57704 M-250 Date Rec'd: 06/11/2009 P.O. # RP2715

Indicator Cells Inoculated: Date/Initials: 6/11/09 / JA

Fixation: Date/Initials: 6/15/09 / JA

Staining: Date/Initials: 6/15/09 / JA

TEST/CONTROL ARTICLE:

TE06-FTDL-01-K #4873

LOT# NA

Wicell QA

DNA FLUROCHROME ASSAY RESULTS:

X **NEGATIVE:** A reaction with staining limited to the nuclear region, which indicates no mycoplasmal contamination.

 POSITIVE: A significant amount of extranuclear staining which strongly suggests mycoplasmal contamination.

 INCONCLUSIVE:
 A significant amount of extranuclear staining consistent with low - level mycoplasmal contamination or nuclear degeneration.

 A significant amount of extranuclear staining consistent with bacterial, fungal or other microbial contaminant or viral CPE. Morphology not consistent for mycoplasmal contamination.

COMMENTS:

Date: 6/15/09 Results Read by: JA Date of Review: 6/15/09 Reviewed by: u

Report Date: June 10, 2009

Case Details:

Cell Line: TE06-FTDL-01 (4873)

Passage #: 49

Date Completed: 6/10/2009

Cell Line Gender: Male

Investigator: National Stem Cell Bank

Specimen: hESC on MEF feeder

Date of Sample: 6/3/2009


Tests, Reason for: FTDL

Results: 46,XY

Completed by _____, MS, CLSp(CG), on 6/9/2009

Reviewed and interpreted by _____, PhD, FACMG, on 6/10/2009

Interpretation: No abnormalities were detected at the stated band level of resolution.



Cell: S01-01

Slide: A

Slide Type: Karyotyping

Cell Results: Karyotype: 46,XY

of Cells Counted: 20

of Cells Karyotyped: 4

of Cells Analyzed: 8

Band Level: 475-575

Results Transmitted by Fax / Email / Post

Sent By: _____

QC Review By: _____

Date: _____

Sent To: _____

Results Recorded: _____